## Corrigendum

Due to some technical problem of software while in converting file in PDF some mistakes have been noticed in the tender No: 14-VI/MS(882)2019PB/T-02, dated 03 June, 2019 under tender reference 2019\_CSIR\_443784\_1. Hence corrigendum against Sec. 1.3 (iii), (iv), (v) and at Sec. 1.11 (vi) is issued as under. Rest of the parameters of our tenders are correct.

1.3	Photodetectors: Point (i) & (ii) are correct.	
(iii)	Spectral match of detector $f_i$	f, M1.5% including the spectral influence of the mirror(s) Supplier should submit calibration report issued byISO/IEC 17025:2005 or latest, accredited metrology institute participant in CIPM-MRA Multipoint calibration in the wavelengths range of the V() function
(iv)	Accuracy class of photometer <b>f</b> <sub>total</sub>	<ul> <li> f<sub>total</sub>≤ 3.0% including influence of the mirror (for the perpendicular light incidence Cosine response f₂=0 and the maximum value for f<sub>total</sub> is 3%, as per the statement at point no.3, below the table 3 of EN 13032-01:2004) (As per DIN 5032 Part 7 Class L, EN 13032-01:2004 Table 3)</li> <li>Supplier should submit value of f<sub>total</sub>, as per EN 13032-01:2004 Table 3, quoting each contributing component values for f<sub>total</sub>, at the time of submission of tender documents</li> <li>Supplier quoting lower values for any component(s) of f<sub>total</sub> than those nominal values stated in EN 13032-01 Table 3, should submit calibration report in support of uncertainty claim for the components, issued byISO/IEC 17025:2005 or latest accredited metrology institute participant in CIPM-MRA at the time of submission of tender documents</li> </ul>
(v)	Spectral mismatch correction factor $f^*$	fm1.0% including spectral influence of mirror as per CIE S-025/E:2015 Annex. C.3.5
		Supplier should submit a test report issued byISO/IEC 17025:2005 or latest accredited metrology institute participant in CIPM-MRA Factor correction by means of software correction or by employment of onboard spectrometer should not be incorporated

1.11	System Calibration and verification of performance ( Point i-v are correct	
(vi)	Calibration of the relative spectral irradiance	-Calibration of the relative spectral irradiance responsivity of the photometer head(s) and mirror assembly
	responsivity of the photometer heads and	-Verification of the $f_I$ quality index of the photometer head and mirror assembly for the V( ) correction according to EN 13032-1
	mirror	- Verification of the spectral mismatch indices of the photometer head and mirror assembly as required with specific spectra in accordance of CIE 127:2007 Fig. 2
		Certificate to be issued by a National Metrology Institute

Stores & Purchase Officer
CSIR-National Physical Laboratory,
New Delhi