



सी.एस.आई.आर. राष्ट्रीय भौतिक प्रयोगशाला
CSIR - NATIONAL PHYSICAL LABORATORY
(वैज्ञानिक तथा औद्योगिक अनुसंधान परिषद)
(Council of Scientific & Industrial Research)

डा० के.एस. कृष्णन् मार्ग, नई दिल्ली - 110012, (भारत)
Dr. K. S. Krishnan Marg, New Delhi - 110012, (INDIA)



Contact: 011-2584-0038

011-2584-0019

Email: spo.nplindia@csir.res.in
purchasesol.nplindia@csir.res.in

From: Director, CSIR-NPL

No. 14-VIII/VE(54-PAC)26-PB/T-40

Dated: 10.09.2026

CORRIGENDUM

With reference to NPL's Global Tender ID: **2026_CSIR_289806_1** for "**Balance of 40 Kg Capacity**".

It is hereby informed that the above-mentioned PAC tender has been published on the Central Public Procurement Portal (CPP Portal).

However, due to an inadvertent omission, the technical specification of the item has not been uploaded along with the tender document.

The **detailed technical specification (Annexure-A)** is hereby being published shall be treated as an integral part of the tender document.

Please also note that bids submitted without taking these changes into consideration will be rejected summarily

All other terms & conditions of said tender will remain the same. The same is also available on CSIR-NPL official website <http://www.nplindia.in> under Tender link.


Stores & Purchase Officer

ANNEXURE-A

Balance of 40Kg Capacity

S.NO.	Technical Specifications
1.	Weighing Capacity: 40 Kg (min) or higher
2.	Readability: 1 mg
3.	Repeatability: 2 mg
4.	Linearity: 20 mg
5.	Eccentric load deviation: 3.5 mg mm
6.	Stabilization time: 5sec
7.	Power Supply: 230V AC/ 50 Hz
8.	Interfaces: RS23C/ USB/LAN
9.	Weighing Pan Dimension (W ×D): 400×300 mm
10.	Anti-Vibration graphite plate/table for placing the balance
11.	Warranty: 12 months after installation and commission
12.	All the cables, required spares should be provided with the system
13.	Installation and demonstration should be done on site, as per the specification
14.	Acceptance Test Post Installation: Successful demonstration of working of the instrument at CSIR-NPL